



WASHINGTON STATE PATROL SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

West Side

DATE: June 22, 2005 INITIALS: DEC OWNER: WSP
 DISTRICT: 2 ADDRESS: I-5 North MP 141
 LOCATION: Sentag No 26 COUNTY: King
 MAKE: Rice Lake TYPE: truck scale ACCURACY CLASS: III
 MODEL: IQ 310 ID. NO: 98438345 CAPACITY: 100,000 lbs SECTIONS: 2
 MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: electronic PLATFORM: 10' x 12'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<i>Center</i>	<i>4000</i>	<i>4000</i>	<i>0</i>	<i>Sec 1</i>	<i>20,000</i>	<i>20,000</i>	<i>0</i>
<i>"</i>	<i>8000</i>	<i>8000</i>	<i>0</i>				
<i>"</i>	<i>12000</i>	<i>12000</i>	<i>0</i>				
<i>"</i>	<i>16000</i>	<i>16000</i>	<i>0</i>	<i>Sec 2</i>	<i>20,000</i>	<i>20,000</i>	<i>0</i>
<i>Center</i>	<i>20,000</i>	<i>20,000</i>	<i>0</i>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 2205-7 and 8


I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Cronin

Date:

June 22, 2005




State of Washington
 County of Thurston
 Signed and sworn to (or affirmed) before
 me on June 22, 2005 by

Michael J. Evans
 (Signature of Notary Public)

Printed Name of Notary

My appointment expires

12-27-07
je



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS